

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
Data File : VL032552.D
Acq On : 27 Sep 2018 16:29
Operator : SY/AP
Sample : QC091218 CAN 10282
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC091218 CAN 10282

Quant Time: Sep 29 22:59:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Sep 25 05:17:15 2018
QLast Update : Tue Sep 25 05:17:15 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1386345	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	3163633	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	3258648	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.61	95	2251953	9.16	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	91.60%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
Data File : VL032552.D
Acq On : 27 Sep 2018 16:29
Operator : SY/AP
Sample : QC091218 CAN 10282
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC091218 CAN 10282

Quant Time: Sep 29 22:59:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep 25 05:17:15 2018
QLast Update : Tue Sep 25 05:17:15 2018
Response via : Initial Calibration

