

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
 Data File : VL032561.D
 Acq On : 27 Sep 2018 22:44
 Operator : SY/AP
 Sample : J5034-07DL2 40X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVE-INF1-092018DL2

Manual Integrations
 APPROVED

MMDadoda
 10/1/2018 11:57:44 AM

Quant Time: Sep 28 10:29:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Sep 28 10:29:20 2018
 QLast Update : Tue Sep 25 05:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1313460	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	2866631	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	3065557	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.61	95	2148911	9.30	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
7) Chloroethane	3.62	64	8342m	0.364	ppbv	
12) 1,1,2-Trichlorotrifluoroet	4.58	101	62718	0.531	ppbv	98
24) 1,1-Dichloroethane	5.11	63	555860	3.276	ppbv	99
31) cis-1,2-Dichloroethene	5.66	61	171832	1.428	ppbv	96
32) 1,1,1-Trichloroethane	6.65	97	655838	3.174	ppbv	99
38) Trichloroethene	7.99	130	341946	3.856	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
Data File : VL032561.D
Acq On : 27 Sep 2018 22:44
Operator : SY/AP
Sample : J5034-07DL2 40X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVE-INF1-092018DL2

Manual Integrations
APPROVED

MMDadoda
10/1/2018 11:57:44 AM

Quant Time: Sep 28 10:29:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep
QLast Update : Tue Sep 25 05:17:15 2018
Response via : Initial Calibration

