

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
 Data File : VL032577.D
 Acq On : 28 Sep 2018 10:33
 Operator : SY/AP
 Sample : J5094-09 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CSS-1

Manual Integrations
 APPROVED

MMDadoda
 10/1/2018 11:58:04 AM

Quant Time: Sep 28 23:50:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep 28 2018
 QLast Update : Tue Sep 25 05:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1274369	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	2780468	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	2791718	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.61	95	2138978	10.16	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chlorodifluoromethane	3.03	51	58623	0.553	ppbv	99
13) Ethanol	3.68	45	125816	9.103	ppbv	96
15) Acetone	3.93	43	69942	0.625	ppbv	96
19) Isopropyl Alcohol	4.08	45	9641m	0.124	ppbv	
20) Methylene Chloride	4.44	84	34161	0.631	ppbv #	81
38) Trichloroethene	7.99	130	16752m	0.195	ppbv	
52) Tetrachloroethene	11.42	164	4503844	57.566	ppbv	99
59) m/p-Xylene	13.16	91	103907m	0.345	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
Data File : VL032577.D
Acq On : 28 Sep 2018 10:33
Operator : SY/AP
Sample : J5094-09 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CSS-1

Manual Integrations
APPROVED

MMDadoda
10/1/2018 11:58:04 AM

Quant Time: Sep 28 23:50:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep
QLast Update : Tue Sep 25 05:17:15 2018
Response via : Initial Calibration

