

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
 Data File : VL032578.D
 Acq On : 28 Sep 2018 11:14
 Operator : SY/AP
 Sample : J5094-10 10X
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 PSS-1

Manual Integrations
 APPROVED

MMDadoda
 10/1/2018 11:58:05 AM

Quant Time: Sep 29 22:46:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep 2
 QLast Update : Tue Sep 25 05:17:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1259909	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	2847188	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	3042727	10.00	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.61	95	2154313	9.39	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chlorodifluoromethane	3.03	51	135360	1.292	ppbv	98
13) Ethanol	3.68	45	70350	5.148	ppbv	98
15) Acetone	3.93	43	241323m	2.182	ppbv	
19) Isopropyl Alcohol	4.07	45	11138m	0.145	ppbv	
26) Hexane	5.80	57	12674m	0.105	ppbv	
29) Chloroform	5.87	83	179776	0.967	ppbv	95
52) Tetrachloroethene	11.41	164	3609473	45.053	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092718\
Data File : VL032578.D
Acq On : 28 Sep 2018 11:14
Operator : SY/AP
Sample : J5094-10 10X
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
PSS-1

Manual Integrations
APPROVED

MMDadoda
10/1/2018 11:58:05 AM

Quant Time: Sep 29 22:46:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Sep
QLast Update : Tue Sep 25 05:17:15 2018
Response via : Initial Calibration

