

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092721\
 Data File : VL037684.D
 Acq On : 27 Sep 2021 17:21
 Operator : SY/AP
 Sample : M3896-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MIDDLE

Manual Integrations
 APPROVED

MMDadoda
 10/1/2021 11:05:06 AM

Quant Time: Sep 30 10:34:16 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 1
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1645427	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.18	114	4641875	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.10	117	4076256	10.00	ppbv	0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.42	95	2924797	9.93	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Ethanol	3.63	45	22523m	2.103	ppbv	
15) Acetone	3.87	43	380013	2.223	ppbv	89
19) Isopropyl Alcohol	4.00	45	90665m	0.975	ppbv	
32) 1,1,1-Trichloroethane	6.52	97	70182	0.262	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092721\
Data File : VL037684.D
Acq On : 27 Sep 2021 17:21
Operator : SY/AP
Sample : M3896-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MIDDLE

Quant Time: Sep 30 10:34:16 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
10/1/2021 11:05:06 AM

