

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL092821\
 Data File : VL037708.D
 Acq On : 28 Sep 2021 18:46
 Operator : SY/AP
 Sample : M3948-01DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 09/29/2021
 Supervised By : Amit Patel 10/01/2021

Quant Time: Sep 29 20:46:36 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1289401	10.00	ppbv	0.03
33) 1,4-Difluorobenzene	7.19	114	4018528	10.00	ppbv	0.03
55) Chlorobenzene-d5	12.10	117	3418666	10.00	ppbv	0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2443197	9.89	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.90%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
13) Ethanol	3.62	45	15743m	1.876	ppbv	
15) Acetone	3.86	43	2446770	18.264	ppbv	100
29) Chloroform	5.76	83	67666m	0.301	ppbv	
34) 2-Butanone	5.27	43	165551m	1.014	ppbv	
52) Tetrachloroethene	11.23	164	30407m	0.236	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL092821\
Data File : VL037708.D
Acq On : 28 Sep 2021 18:46
Operator : SY/AP
Sample : M3948-01DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Quant Time: Sep 29 20:46:36 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL091621
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 09/29/2021
Supervised By :Amit Patel 10/01/2021

