

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_L\DATA\VL092821\  
 Data File : VL037699.D  
 Acq On : 28 Sep 2021 11:34  
 Operator : SY/AP  
 Sample : VSTDCCC010  
 Misc : 400mL/MSVOA\_L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 LabSampleId :  
 VSTDCCC010

Quant Time: Sep 29 03:46:19 2021  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_L\METHODS\VL091621AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA\_L Thu Sep 16 13:42:46 2021  
 QLast Update : Thu Sep 16 13:42:46 2021  
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I  | Bromochloromethane          | 1.000 | 1.000 | 0.0   | 85    | 0.02     |
| 2 T  | Dichlorodifluoromethane     | 1.810 | 1.278 | 29.4  | 56    | 0.02     |
| 3    | Chlorodifluoromethane       | 1.886 | 1.801 | 4.5   | 89    | 0.02     |
| 4    | Chloromethane               | 0.697 | 0.684 | 1.9   | 89    | 0.02     |
| 5 T  | Vinyl Chloride              | 0.655 | 0.677 | -3.4  | 87    | 0.02     |
| 6 T  | Bromomethane                | 0.380 | 0.349 | 8.2   | 80    | 0.01     |
| 7    | Chloroethane                | 0.245 | 0.239 | 2.4   | 93    | 0.02     |
| 8 T  | Dichlorotetrafluoroethane   | 1.561 | 1.536 | 1.6   | 88    | 0.02     |
| 9 T  | Propene                     | 0.665 | 0.676 | -1.7  | 95    | 0.02     |
| 10 T | Heptane                     | 1.553 | 1.684 | -8.4  | 91    | 0.02     |
| 11 T | Trichlorofluoromethane      | 1.453 | 1.432 | 1.4   | 90    | 0.02     |
| 12 T | 1,1,2-Trichlorotrifluoroeth | 1.072 | 1.063 | 0.8   | 92    | 0.02     |
| 13   | Ethanol                     | 0.065 | 0.059 | 9.2   | 84    | 0.02     |
| 14 T | Bromoethene                 | 0.488 | 0.477 | 2.3   | 89    | 0.02     |
| 15 T | Acetone                     | 1.039 | 0.818 | 21.3  | 87    | 0.02     |
| 16 T | 1,3-Butadiene               | 0.649 | 0.680 | -4.8  | 92    | 0.02     |
| 17   | tert-Butyl alcohol          | 0.881 | 0.946 | -7.4  | 90    | 0.02     |
| 18 T | 1,1-Dichloroethene          | 0.500 | 0.485 | 3.0   | 92    | 0.02     |
| 19 T | Isopropyl Alcohol           | 0.565 | 0.567 | -0.4  | 91    | 0.02     |
| 20 T | Methylene Chloride          | 0.496 | 0.428 | 13.7  | 91    | 0.02     |
| 21 T | Allyl Chloride              | 0.776 | 0.790 | -1.8  | 91    | 0.02     |
| 22 T | trans-1,2-Dichloroethene    | 0.503 | 0.494 | 1.8   | 90    | 0.02     |
| 23 T | Vinyl Acetate               | 1.438 | 1.562 | -8.6  | 91    | 0.02     |
| 24 T | 1,1-Dichloroethane          | 1.030 | 0.982 | 4.7   | 89    | 0.02     |
| 25 T | Ethyl Acetate               | 2.516 | 2.604 | -3.5  | 90    | 0.02     |
| 26 T | Hexane                      | 1.195 | 1.245 | -4.2  | 90    | 0.02     |
| 27 T | Carbon Disulfide            | 1.094 | 1.171 | -7.0  | 90    | 0.02     |
| 28 T | Methyl tert-Butyl Ether     | 1.053 | 1.199 | -13.9 | 102   | 0.02     |
| 29 T | Chloroform                  | 1.745 | 1.693 | 3.0   | 87    | 0.02     |
| 30 T | Cyclohexane                 | 1.022 | 1.047 | -2.4  | 91    | 0.03     |
| 31 T | cis-1,2-Dichloroethene      | 1.085 | 1.167 | -7.6  | 91    | 0.02     |
| 32 T | 1,1,1-Trichloroethane       | 1.625 | 1.608 | 1.0   | 86    | 0.02     |
| 33 I | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 89    | 0.03     |
| 34 T | 2-Butanone                  | 0.406 | 0.376 | 7.4   | 89    | 0.02     |
| 35 T | Carbon Tetrachloride        | 0.566 | 0.549 | 3.0   | 86    | 0.03     |
| 36 T | Benzene                     | 0.834 | 0.841 | -0.8  | 91    | 0.02     |
| 37 T | 1,2-Dichloroethane          | 0.410 | 0.396 | 3.4   | 87    | 0.02     |
| 38 T | Trichloroethene             | 0.348 | 0.326 | 6.3   | 88    | 0.03     |
| 39 T | 1,2-Dichloropropane         | 0.319 | 0.319 | 0.0   | 89    | 0.02     |
| 40 T | 1,4-Dioxane                 | 0.078 | 0.079 | -1.3  | 85    | 0.03     |
| 41 T | Tetrahydrofuran             | 0.191 | 0.198 | -3.7  | 89    | 0.03     |
| 42 T | Bromodichloromethane        | 0.588 | 0.593 | -0.9  | 87    | 0.02     |
| 43   | Methyl Methacrylate         | 0.277 | 0.301 | -8.7  | 89    | 0.02     |
| 44 T | 2,2,4-Trimethylpentane      | 1.402 | 1.386 | 1.1   | 90    | 0.03     |
| 45 T | t-1,3-Dichloropropene       | 0.218 | 0.224 | -2.8  | 76    | 0.03     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_L\DATA\VL092821\  
 Data File : VL037699.D  
 Acq On : 28 Sep 2021 11:34  
 Operator : SY/AP  
 Sample : VSTDCCC010  
 Misc : 400mL/MSVOA\_L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 LabSampleId :  
 VSTDCCC010

Quant Time: Sep 29 03:46:19 2021  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_L\METHODS\VL091621AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Thu Sep 16 13:42:46 2021  
 QLast Update : Thu Sep 16 13:42:46 2021  
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 150%

|      | Compound                  | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|------|---------------------------|-------|-------|--------|-------|----------|
| 46 T | cis-1,3-Dichloropropene   | 0.294 | 0.304 | -3.4   | 79    | 0.02     |
| 47 T | 1,1,2-Trichloroethane     | 0.353 | 0.335 | 5.1    | 86    | 0.03     |
| 48 T | Dibromochloromethane      | 0.493 | 0.520 | -5.5   | 86    | 0.03     |
| 49 T | Bromoform                 | 0.394 | 0.422 | -7.1   | 81    | 0.03     |
| 50 T | 4-Methyl-2-Pentanone      | 0.516 | 0.561 | -8.7   | 93    | 0.02     |
| 51 T | 2-Hexanone                | 0.228 | 0.270 | -18.4  | 95    | 0.03     |
| 52 T | Tetrachloroethene         | 0.320 | 0.313 | 2.2    | 90    | 0.03     |
| 53 T | Toluene                   | 0.927 | 0.970 | -4.6   | 90    | 0.03     |
| 54 T | 1,2-Dibromoethane         | 0.484 | 0.470 | 2.9    | 85    | 0.03     |
| 55 I | Chlorobenzene-d5          | 1.000 | 1.000 | 0.0    | 91    | 0.03     |
| 56   | 1,1,1,2-Tetrachloroethane | 0.415 | 0.381 | 8.2    | 84    | 0.03     |
| 57 T | Chlorobenzene             | 0.831 | 0.737 | 11.3   | 84    | 0.03     |
| 58 T | Ethyl Benzene             | 1.327 | 1.313 | 1.1    | 87    | 0.03     |
| 59 T | m/p-Xylene                | 1.157 | 1.079 | 6.7    | 86    | 0.03     |
| 60 T | o-Xylene                  | 1.101 | 1.020 | 7.4    | 86    | 0.03     |
| 61 T | Styrene                   | 0.543 | 0.575 | -5.9   | 84    | 0.03     |
| 62   | Isopropylbenzene          | 1.538 | 1.418 | 7.8    | 86    | 0.03     |
| 63 T | 1,1,2,2-Tetrachloroethane | 0.862 | 0.724 | 16.0   | 86    | 0.03     |
| 64   | n-propylbenzene           | 0.397 | 0.380 | 4.3    | 85    | 0.03     |
| 65   | tert-Butylbenzene         | 1.387 | 1.242 | 10.5   | 87    | 0.03     |
| 66 T | Benzyl Chloride           | 0.076 | 0.068 | 10.5   | 76    | 0.03     |
| 67   | sec-Butylbenzene          | 1.882 | 1.724 | 8.4    | 87    | 0.03     |
| 68 S | 1-Bromo-4-Fluorobenzene   | 0.723 | 0.723 | 0.0    | 86    | 0.03     |
| 69   | p-Isopropyltoluene        | 1.537 | 1.441 | 6.2    | 86    | 0.03     |
| 70   | n-Butylbenzene            | 1.454 | 1.400 | 3.7    | 86    | 0.03     |
| 71   | 2-Chlorotoluene           | 1.102 | 1.038 | 5.8    | 84    | 0.03     |
| 72 T | 4-Ethyltoluene            | 1.264 | 1.212 | 4.1    | 85    | 0.03     |
| 73 T | 1,3,5-Trimethylbenzene    | 1.178 | 1.092 | 7.3    | 85    | 0.03     |
| 74 T | 1,2,4-Trimethylbenzene    | 1.241 | 1.132 | 8.8    | 86    | 0.03     |
| 75 T | 1,3-Dichlorobenzene       | 0.771 | 0.695 | 9.9    | 85    | 0.03     |
| 76 T | 1,4-Dichlorobenzene       | 0.760 | 0.674 | 11.3   | 83    | 0.03     |
| 77 T | 1,2-Dichlorobenzene       | 0.731 | 0.663 | 9.3    | 86    | 0.03     |
| 78 T | Hexachloro-1,3-Butadiene  | 0.615 | 0.493 | 19.8   | 89    | 0.03     |
| 79 T | Naphthalene               | 0.762 | 0.827 | -8.5   | 90    | 0.04     |
| 80 T | Naphthalene,2-methyl-     | 0.174 | 0.231 | -32.8# | 91    | 0.02     |
| 81 T | 1,2,4-Trichlorobenzene    | 0.489 | 0.507 | -3.7   | 92    | 0.04     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0