

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092921\
 Data File : VL037734.D
 Acq On : 30 Sep 2021 9:20
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

SAM
 10/1/2021 11:09:20 AM

Quant Time: Oct 01 01:16:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
 QLast Update : Thu Sep 16 13:42:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.658	49	1150649	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.163	114	3519100	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.072	117	3009273	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.397	95	2119061	9.74	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.40%
Target Compounds						
						Qvalue
9) Propene	3.018	41	17419m	0.23	ppbv	
15) Acetone	3.857	43	68531m	0.57	ppbv	
20) Methylene Chloride	4.343	84	27773m	0.49	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL092921\
Data File : VL037734.D
Acq On : 30 Sep 2021 9:20
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

SAM
10/1/2021 11:09:20 AM

Quant Time: Oct 01 01:16:43 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL091621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Sep 16 13:42:46 2021
QLast Update : Thu Sep 16 13:42:46 2021
Response via : Initial Calibration

