

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL093025\
 Data File : VL043023.D
 Acq On : 30 Sep 2025 11:53
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/01/2025
 Supervised By :Mahesh Dadoda 10/03/2025

Quant Time: Sep 30 23:58:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL093025AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Sep 30 23:54:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	142955	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	436249	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	383741	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	281670	9.689	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.900%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	363	0.040	ppbv	100
32) 1,1,1-Trichloroethane	3.369	97	1171	0.034	ppbv	99
35) Carbon Tetrachloride	3.758	117	1365m	0.039	ppbv	
38) Trichloroethene	4.538	130	690m	0.031	ppbv	
52) Tetrachloroethene	8.231	164	698m	0.040	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	843m	0.030	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL093025\
Data File : VL043023.D
Acq On : 30 Sep 2025 11:53
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/01/2025
Supervised By :Mahesh Dadoda 10/03/2025

Quant Time: Sep 30 23:58:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL093025AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Sep 30 23:54:02 2025
Response via : Initial Calibration

