

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL100318\
 Data File : VL032590.D
 Acq On : 3 Oct 2018 21:57
 Operator : SY/AP
 Sample : J5094-11DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 KSS-1DL

Quant Time: Oct 04 10:00:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Oct 01 06:49:26 2018
 QLast Update : Mon Oct 01 06:49:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.79	49	1029896	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.32	114	2596183	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.27	117	2804657	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.60	95	1989860	9.41	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.03	51	1118123	13.055	ppbv	97
13) Ethanol	3.68	45	3442452	308.196	ppbv	96
15) Acetone	3.92	43	545171	6.031	ppbv	100
19) Isopropyl Alcohol	4.05	45	111214	1.766	ppbv #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL100318\
Data File : VL032590.D
Acq On : 3 Oct 2018 21:57
Operator : SY/AP
Sample : J5094-11DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
KSS-1DL

Quant Time: Oct 04 10:00:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL092418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 01 06:49:26 2018
QLast Update : Mon Oct 01 06:49:26 2018
Response via : Initial Calibration

