

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100421\
 Data File : VL037763.D
 Acq On : 4 Oct 2021 15:57
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/05/2021
 Supervised By :Semsettin Yesilyurt 10/05/2021

Quant Time: Oct 04 17:32:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Oct 04 17:32:43 2021
 QLast Update : Mon Oct 04 12:52:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	1402853	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.176	114	4111572	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.085	117	3539703	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.413	95	2461157	9.62	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.20%
Target Compounds						
5) Vinyl Chloride	3.259	62	2473m	0.03	ppbv	Qvalue
32) 1,1,1-Trichloroethane	6.513	97	7057m	0.03	ppbv	
35) Carbon Tetrachloride	7.018	117	7775m	0.03	ppbv	
38) Trichloroethene	7.835	130	4109m	0.03	ppbv	
52) Tetrachloroethene	11.224	164	5068m	0.04	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.680	83	11813m	0.04	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100421\
Data File : VL037763.D
Acq On : 4 Oct 2021 15:57
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Oct 04 17:32:43 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 04 12:52:19 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 10/05/2021
Supervised By : Semsettin Yesilyurt 10/05/2021

