

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
 Data File : VL037809.D
 Acq On : 7 Oct 2021 23:40
 Operator : SY/AP
 Sample : M4079-14DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P08DL

Manual Integrations
 APPROVED

apatel
 10/14/2021 1:19:26 PM

Quant Time: Oct 08 11:08:18 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 0
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1345914	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	4109344	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	3509658	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2411868	9.83	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.01	41	15749m	0.166	ppbv	
13) Ethanol	3.61	45	29988m	3.854	ppbv	
15) Acetone	3.86	43	61599m	0.510	ppbv	
20) Methylene Chloride	4.33	84	37479m	0.614	ppbv	
26) Hexane	5.67	57	24176m	0.145	ppbv	
29) Chloroform	5.73	83	33866m	0.142	ppbv	
52) Tetrachloroethene	11.20	164	264743m	1.939	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
Data File : VL037809.D
Acq On : 7 Oct 2021 23:40
Operator : SY/AP
Sample : M4079-14DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P08DL

Manual Integrations
APPROVED

apatel
10/14/2021 1:19:26 PM

Quant Time: Oct 08 11:08:18 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

