

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
 Data File : VL037811.D
 Acq On : 8 Oct 2021 00:59
 Operator : SY/AP
 Sample : M4079-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P45DL

Manual Integrations
 APPROVED

apatel
 10/14/2021 1:19:23 PM

Quant Time: Oct 08 11:13:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 0
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1437095	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	4155620	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	3540346	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2505506	10.12	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.02	41	16622m	0.164	ppbv	
11) Trichlorofluoromethane	3.94	101	109215	0.533	ppbv	90
13) Ethanol	3.59	45	1950631	234.763	ppbv	80
15) Acetone	3.84	43	215523	1.671	ppbv	95
19) Isopropyl Alcohol	3.95	45	128122	1.596	ppbv #	92
20) Methylene Chloride	4.33	84	99229	1.522	ppbv #	92
26) Hexane	5.67	57	54710	0.307	ppbv #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
Data File : VL037811.D
Acq On : 8 Oct 2021 00:59
Operator : SY/AP
Sample : M4079-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P45DL

Manual Integrations
APPROVED

apatel
10/14/2021 1:19:23 PM

Quant Time: Oct 08 11:13:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

