

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
 Data File : VL037813.D
 Acq On : 8 Oct 2021 2:17
 Operator : SY/AP
 Sample : M4079-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P46DL

Quant Time: Oct 08 11:19:48 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1448145	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.16	114	4003806	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	3542616	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2393090	9.66	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.01	41	13394	0.131	ppbv	95
11) Trichlorofluoromethane	3.94	101	82604	0.400	ppbv	98
13) Ethanol	3.59	45	1842902	220.105	ppbv	81
15) Acetone	3.84	43	226151	1.740	ppbv	95
19) Isopropyl Alcohol	3.95	45	131955	1.632	ppbv #	87
20) Methylene Chloride	4.34	84	39714	0.605	ppbv #	80
26) Hexane	5.68	57	25552m	0.142	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
Data File : VL037813.D
Acq On : 8 Oct 2021 2:17
Operator : SY/AP
Sample : M4079-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P46DL

Quant Time: Oct 08 11:19:48 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

