

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
 Data File : VL037819.D
 Acq On : 8 Oct 2021 6:14
 Operator : SY/AP
 Sample : M4079-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P13DL

Manual Integrations
 APPROVED

apatel
 10/14/2021 1:19:14 PM

Quant Time: Oct 08 11:33:39 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 0
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1451186	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4115242	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.07	117	3602202	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2517792	10.00	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.01	41	16398m	0.160	ppbv	
13) Ethanol	3.63	45	10830m	1.291	ppbv	
15) Acetone	3.86	43	80687	0.619	ppbv #	82
20) Methylene Chloride	4.33	84	33562	0.510	ppbv #	80
52) Tetrachloroethene	11.21	164	136314	0.997	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
Data File : VL037819.D
Acq On : 8 Oct 2021 6:14
Operator : SY/AP
Sample : M4079-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P13DL

Manual Integrations
APPROVED

apatel
10/14/2021 1:19:14 PM

Quant Time: Oct 08 11:33:39 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

