

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
 Data File : VL037821.D
 Acq On : 8 Oct 2021 7:32
 Operator : SY/AP
 Sample : M4079-12DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0P12DL

Manual Integrations
 APPROVED

apatel
 10/14/2021 1:19:11 PM

Quant Time: Oct 08 11:38:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Oct 0
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1375633	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.16	114	3919582	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.07	117	3430617	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2411549	10.05	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.01	41	12185m	0.125	ppbv	
13) Ethanol	3.62	45	9250m	1.163	ppbv	
15) Acetone	3.86	43	85343m	0.691	ppbv	
20) Methylene Chloride	4.33	84	23110m	0.370	ppbv	
26) Hexane	5.67	57	19815m	0.116	ppbv	
52) Tetrachloroethene	11.21	164	13448m	0.103	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
Data File : VL037821.D
Acq On : 8 Oct 2021 7:32
Operator : SY/AP
Sample : M4079-12DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0P12DL

Manual Integrations
APPROVED

apatel
10/14/2021 1:19:11 PM

Quant Time: Oct 08 11:38:37 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

