

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
 Data File : VL037824.D
 Acq On : 8 Oct 2021 9:42
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

apatel
 10/14/2021 1:19:07 PM

Quant Time: Oct 09 18:11:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Oct 0
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1277054	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	3772060	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	3117534	10.00	ppbv	-0.02

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.39	95	2244789	10.30	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.00%

Target Compounds					Qvalue
15) Acetone	3.87	43	35404m	0.309	ppbv
20) Methylene Chloride	4.34	84	18798m	0.325	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL100721\
Data File : VL037824.D
Acq On : 8 Oct 2021 9:42
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

apatel
10/14/2021 1:19:07 PM

Quant Time: Oct 09 18:11:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL100421AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

