

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100825\
 Data File : VL043071.D
 Acq On : 08 Oct 2025 22:43
 Operator : SY/MD
 Sample : Q3320-13DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Quant Time: Oct 09 01:25:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	165692	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	425630	10.000	ppbv	0.02
55) Chlorobenzene-d5	8.875	117	331771	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	229192	9.590	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.900%
Target Compounds						
					Qvalue	
15) Acetone	1.835	43	9361	0.403	ppbv #	89
20) Methylene Chloride	2.062	84	2814	0.278	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100825\
Data File : VL043071.D
Acq On : 08 Oct 2025 22:43
Operator : SY/MD
Sample : Q3320-13DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Quant Time: Oct 09 01:25:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Oct 03 08:23:34 2025
Response via : Initial Calibration

