

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100825\
 Data File : VL043077.D
 Acq On : 09 Oct 2025 02:01
 Operator : SY/MD
 Sample : Q3320-07DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA4DL

Quant Time: Oct 09 02:49:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Oct 03 08:23:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	166091	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	408901	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.872	117	318833	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	217373	9.464	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.600%
Target Compounds						
					Qvalue	
9) Propene	1.486	41	1209m	0.111	ppbv	
13) Ethanol	1.719	45	8343	19.816	ppbv #	31
15) Acetone	1.835	43	11089	0.477	ppbv	100
20) Methylene Chloride	2.062	84	3072	0.303	ppbv	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100825\
Data File : VL043077.D
Acq On : 09 Oct 2025 02:01
Operator : SY/MD
Sample : Q3320-07DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA4DL

Quant Time: Oct 09 02:49:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Oct 03 08:23:34 2025
Response via : Initial Calibration

