

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101725\
 Data File : VL043135.D
 Acq On : 17 Oct 2025 11:03
 Operator : SY/MD
 Sample : QC101525 CAN 10332
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC101525 CAN 10332

Quant Time: Oct 18 04:34:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Oct 15 02:12:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.793	49	86103	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	266279	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	226940	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.380	95	184716	10.472	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL101725\
Data File : VL043135.D
Acq On : 17 Oct 2025 11:03
Operator : SY/MD
Sample : QC101525 CAN 10332
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC101525 CAN 10332

Quant Time: Oct 18 04:34:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Oct 15 02:12:58 2025
Response via : Initial Calibration

