

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL101821\  
Data File : VL037866.D  
Acq On : 18 Oct 2021 23:44  
Operator : SY/AP  
Sample : CAN 10333  
Misc : 400mL/MSVOA L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
CAN 10333

Quant Time: Oct 19 12:02:41 2021  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421AIR.M  
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Mon Oct 04 17:43:54 2021  
QLast Update : Mon Oct 04 17:43:54 2021  
Response via : Initial Calibration

| Internal Standards      | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------|-------|------|----------|-------|-------|----------|
| 1) Bromochloromethane   | 5.67  | 49   | 1185017  | 10.00 | ppbv  | 0.00     |
| 33) 1,4-Difluorobenzene | 7.17  | 114  | 3447838  | 10.00 | ppbv  | 0.00     |
| 55) Chlorobenzene-d5    | 12.08 | 117  | 2885864  | 10.00 | ppbv  | 0.00     |

System Monitoring Compounds

|                             |        |       |          |          |      |        |
|-----------------------------|--------|-------|----------|----------|------|--------|
| 68) 1-Bromo-4-Fluorobenzene | 14.41  | 95    | 1992371  | 9.87     | ppbv | 0.00   |
| Spiked Amount               | 10.000 | Range | 65 - 135 | Recovery | =    | 98.70% |

| Target Compounds | Ovalue |
|------------------|--------|
|------------------|--------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL101821\  
Data File : VL037866.D  
Acq On : 18 Oct 2021 23:44  
Operator : SY/AP  
Sample : CAN 10333  
Misc : 400mL/MSVOA L  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_L  
ClientSampleId :  
CAN 10333

Quant Time: Oct 19 12:02:41 2021  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL100421AIR.M  
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Mon Oct 04 17:43:54 2021  
QLast Update : Mon Oct 04 17:43:54 2021  
Response via : Initial Calibration

