

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL102021\
 Data File : VL037878.D
 Acq On : 20 Oct 2021 8:31
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 21 01:38:21 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	1.597	1.169	26.8	80	0.00
3	Chlorodifluoromethane	1.921	2.015	-4.9	96	0.00
4	Chloromethane	0.704	0.759	-7.8	97	0.00
5 T	Vinyl Chloride	0.645	0.724	-12.2	91	0.00
6 T	Bromomethane	0.349	0.373	-6.9	96	0.00
7	Chloroethane	0.237	0.253	-6.8	91	0.00
8 T	Dichlorotetrafluoroethane	1.579	1.641	-3.9	96	0.00
9 T	Propene	0.707	0.713	-0.8	90	0.00
10 T	Heptane	1.606	1.874	-16.7	98	0.00
11 T	Trichlorofluoromethane	1.425	1.507	-5.8	93	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.080	1.182	-9.4	100	0.00
13	Ethanol	0.058	0.064	-10.3	114	0.00
14 T	Bromoethene	0.475	0.492	-3.6	94	0.00
15 T	Acetone	0.898	0.852	5.1	94	0.00
16 T	1,3-Butadiene	0.653	0.696	-6.6	91	0.00
17	tert-Butyl alcohol	0.925	0.995	-7.6	86	0.00
18 T	1,1-Dichloroethene	0.491	0.519	-5.7	91	0.00
19 T	Isopropyl Alcohol	0.558	0.543	2.7	79	0.00
20 T	Methylene Chloride	0.454	0.453	0.2	97	0.00
21 T	Allyl Chloride	0.795	0.824	-3.6	90	0.00
22 T	trans-1,2-Dichloroethene	0.493	0.539	-9.3	93	0.00
23 T	Vinyl Acetate	1.345	1.397	-3.9	92	0.00
24 T	1,1-Dichloroethane	0.979	1.108	-13.2	100	0.00
25 T	Ethyl Acetate	2.519	2.807	-11.4	95	0.00
26 T	Hexane	1.240	1.330	-7.3	94	0.00
27 T	Carbon Disulfide	1.099	1.345	-22.4	97	0.00
28 T	Methyl tert-Butyl Ether	1.033	1.077	-4.3	93	0.00
29 T	Chloroform	1.773	1.903	-7.3	97	0.00
30 T	Cyclohexane	1.052	1.135	-7.9	96	0.00
31 T	cis-1,2-Dichloroethene	1.126	1.226	-8.9	94	0.00
32 T	1,1,1-Trichloroethane	1.681	1.794	-6.7	94	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
34 T	2-Butanone	0.404	0.327	19.1	72	0.00
35 T	Carbon Tetrachloride	0.568	0.641	-12.9	97	0.00
36 T	Benzene	0.854	0.951	-11.4	96	0.00
37 T	1,2-Dichloroethane	0.402	0.444	-10.4	94	0.00
38 T	Trichloroethene	0.337	0.377	-11.9	99	0.00
39 T	1,2-Dichloropropane	0.321	0.361	-12.5	95	0.00
40 T	1,4-Dioxane	0.080	0.066	17.5	76	0.00
41 T	Tetrahydrofuran	0.198	0.222	-12.1	93	0.00
42 T	Bromodichloromethane	0.591	0.685	-15.9	97	0.00
43	Methyl Methacrylate	0.279	0.336	-20.4	95	0.00
44 T	2,2,4-Trimethylpentane	1.435	1.590	-10.8	97	0.00
45 T	t-1,3-Dichloropropene	0.192	0.252	-31.3#	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL102021\
 Data File : VL037878.D
 Acq On : 20 Oct 2021 8:31
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Oct 21 01:38:21 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Oct 04 17:43:54 2021
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.266	0.342	-28.6	95	0.00
47 T	1,1,2-Trichloroethane	0.349	0.384	-10.0	95	0.00
48 T	Dibromochloromethane	0.498	0.597	-19.9	95	0.00
49 T	Bromoform	0.381	0.473	-24.1	92	0.00
50 T	4-Methyl-2-Pentanone	0.526	0.626	-19.0	96	0.00
51 T	2-Hexanone	0.235	0.281	-19.6	90	0.00
52 T	Tetrachloroethene	0.332	0.344	-3.6	96	0.00
53 T	Toluene	0.947	1.097	-15.8	96	0.00
54 T	1,2-Dibromoethane	0.472	0.523	-10.8	94	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
56	1,1,1,2-Tetrachloroethane	0.414	0.468	-13.0	98	0.00
57 T	Chlorobenzene	0.820	0.904	-10.2	98	0.00
58 T	Ethyl Benzene	1.343	1.530	-13.9	95	0.00
59 T	m/p-Xylene	1.153	1.301	-12.8	97	0.00
60 T	o-Xylene	1.105	1.207	-9.2	96	0.00
61 T	Styrene	0.534	0.670	-25.5	94	0.00
62	Isopropylbenzene	1.541	1.726	-12.0	97	0.00
63 T	1,1,2,2-Tetrachloroethane	0.852	0.859	-0.8	95	0.00
64	n-propylbenzene	0.384	0.435	-13.3	95	0.00
65	tert-Butylbenzene	1.356	1.461	-7.7	96	0.00
66 T	Benzyl Chloride	0.067	0.071	-6.0	82	0.00
67	sec-Butylbenzene	1.854	2.018	-8.8	95	0.00
68 S	1-Bromo-4-Fluorobenzene	0.699	0.712	-1.9	82	0.00
69	p-Isopropyltoluene	1.499	1.655	-10.4	95	0.00
70	n-Butylbenzene	1.424	1.599	-12.3	94	0.00
71	2-Chlorotoluene	1.088	1.180	-8.5	92	0.00
72 T	4-Ethyltoluene	1.254	1.477	-17.8	97	0.00
73 T	1,3,5-Trimethylbenzene	1.139	1.320	-15.9	98	0.00
74 T	1,2,4-Trimethylbenzene	1.196	1.335	-11.6	96	0.00
75 T	1,3-Dichlorobenzene	0.741	0.809	-9.2	97	0.00
76 T	1,4-Dichlorobenzene	0.707	0.781	-10.5	95	0.00
77 T	1,2-Dichlorobenzene	0.712	0.764	-7.3	95	0.00
78 T	Hexachloro-1,3-Butadiene	0.581	0.547	5.9	95	0.00
79 T	Naphthalene	0.733	0.873	-19.1	92	0.00
80 T	Naphthalene,2-methyl-	0.197	0.193	2.0	78	0.00
81 T	1,2,4-Trichlorobenzene	0.469	0.531	-13.2	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0