

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102122\
 Data File : VL039751.D
 Acq On : 21 Oct 2022 19:57
 Operator : SY/AP
 Sample : N5259-01 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/24/2022
 Supervised By :Mahesh Dadoda 10/24/2022

Quant Time: Oct 22 04:25:59 2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Sat Oct 22 02:12:34 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	879031	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.651	114	2601422	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2225110	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	1602101	10.143	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.400%
Target Compounds						
13) Ethanol	3.269	45	16279m	4.672	ppbv	Qvalue
15) Acetone	3.494	43	50113m	0.472	ppbv	
19) Isopropyl Alcohol	3.607	45	5248m	0.101	ppbv	
20) Methylene Chloride	3.944	84	30573	0.716	ppbv	93
26) Hexane	5.214	57	33825	0.391	ppbv #	37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102122\
Data File : VL039751.D
Acq On : 21 Oct 2022 19:57
Operator : SY/AP
Sample : N5259-01 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1

Quant Time: Oct 22 04:25:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Sat Oct 22 02:12:34 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/24/2022
Supervised By :Mahesh Dadoda 10/24/2022

