

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102122\
 Data File : VL039752.D
 Acq On : 21 Oct 2022 20:36
 Operator : SY/AP
 Sample : N5259-01DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/24/2022
 Supervised By :Mahesh Dadoda 10/24/2022

Quant Time: Oct 22 04:26:17 2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Sat Oct 22 02:12:34 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	868099	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.651	114	2548618	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2213534	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	1540313	9.803	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.000%
Target Compounds						
13) Ethanol	3.266	45	2935m	0.853	ppbv	Qvalue
15) Acetone	3.504	43	13087	0.125	ppbv #	87
20) Methylene Chloride	3.947	84	17921	0.425	ppbv	94
26) Hexane	5.224	57	8521m	0.100	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL102122\
Data File : VL039752.D
Acq On : 21 Oct 2022 20:36
Operator : SY/AP
Sample : N5259-01DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/24/2022
Supervised By :Mahesh Dadoda 10/24/2022

Quant Time: Oct 22 04:26:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL102122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Sat Oct 22 02:12:34 2022
Response via : Initial Calibration

