

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL102521\
 Data File : VL037897.D
 Acq On : 25 Oct 2021 17:03
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/27/2021
 Supervised By :Mahesh Dadoda 10/27/2021

Quant Time: Oct 26 00:14:57 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Oct 04 17:43:54 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1375096	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	7.19	114	4058433	10.00	ppbv	0.01
55) Chlorobenzene-d5	12.10	117	3323780	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2333064	10.04	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.03	41	12560m	0.129	ppbv	
15) Acetone	3.90	43	15617m	0.127	ppbv	
20) Methylene Chloride	4.36	84	18432m	0.296	ppbv	
32) 1,1,1-Trichloroethane	6.53	97	8549m	0.037	ppbv	
35) Carbon Tetrachloride	7.03	117	8613m	0.037	ppbv	
38) Trichloroethene	7.85	130	5234m	0.038	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL102521\
Data File : VL037897.D
Acq On : 25 Oct 2021 17:03
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Oct 26 00:14:57 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL100421
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Oct 04 17:43:54 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 10/27/2021
Supervised By :Mahesh Dadoda 10/27/2021

