

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110222\
 Data File : VL039814.D
 Acq On : 2 Nov 2022 13:17
 Operator : SY/AP
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/03/2022
 Supervised By :Mahesh Dadoda 11/03/2022

Quant Time: Nov 02 13:55:24 2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Nov 02 11:28:32 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	1103549	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	2983316	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.452	117	2703169	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.761	95	2044111	10.630	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.300%
Target Compounds						
						Qvalue
5) Vinyl Chloride	2.928	62	6296m	0.091	ppbv	
32) 1,1,1-Trichloroethane	5.996	97	14952m	0.098	ppbv	
35) Carbon Tetrachloride	6.484	117	14368m	0.084	ppbv	
38) Trichloroethene	7.279	130	11471m	0.088	ppbv	
52) Tetrachloroethene	10.600	164	8100m	0.094	ppbv	
54) 1,2-Dibromoethane	9.992	107	8908m	0.093	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.037	83	9390m	0.086	ppbv	
79) Naphthalene	21.378	128	17713m	0.095	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110222\
Data File : VL039814.D
Acq On : 2 Nov 2022 13:17
Operator : SY/AP
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022

Quant Time: Nov 02 13:55:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Wed Nov 02 11:28:32 2022
Response via : Initial Calibration

