

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110420\
 Data File : VL035829.D
 Acq On : 4 Nov 2020 10:34
 Operator : SY/AP
 Sample : VL1104ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1104ABL01

Manual Integrations
 APPROVED

MMDadoda
 11/5/2020 2:16:25 PM

Quant Time: Nov 05 03:23:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1015770	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.14	114	4585407	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	4239172	10.00	ppbv	-0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2987888	9.92	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%
Target Compounds						
15) Acetone	3.84	43	22137	0.204	ppbv	Ovalue # 1
20) Methylene Chloride	4.32	84	13414m	0.175	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110420\
Data File : VL035829.D
Acq On : 4 Nov 2020 10:34
Operator : SY/AP
Sample : VL1104ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1104ABL01

Manual Integrations
APPROVED

MMDadoda
11/5/2020 2:16:25 PM

Quant Time: Nov 05 03:23:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

