

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110420\
 Data File : VL035839.D
 Acq On : 4 Nov 2020 18:48
 Operator : SY/AP
 Sample : L4610-05DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV5DL

Manual Integrations
 APPROVED

MMDadoda
 11/5/2020 2:17:04 PM

Quant Time: Nov 05 11:27:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 11:27:39 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1002585	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.14	114	4581190	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	4133103	10.00	ppbv	-0.04

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2932153	9.99	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	123282	2.805	ppbv #	80
15) Acetone	3.83	43	432354	4.034	ppbv	97
20) Methylene Chloride	4.32	84	65568	0.869	ppbv	98
26) Hexane	5.65	57	44954	0.386	ppbv #	43
52) Tetrachloroethene	11.19	164	33011	0.205	ppbv	95
53) Toluene	9.77	91	70615	0.172	ppbv	97
59) m/p-Xylene	12.94	91	71229m	0.167	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110420\
Data File : VL035839.D
Acq On : 4 Nov 2020 18:48
Operator : SY/AP
Sample : L4610-05DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV5DL

Manual Integrations
APPROVED

MMDadoda
11/5/2020 2:17:04 PM

Quant Time: Nov 05 11:27:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

