

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110420\
 Data File : VL035841.D
 Acq On : 4 Nov 2020 20:06
 Operator : SY/AP
 Sample : L4610-06DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV6DDL

Manual Integrations
 APPROVED

MMDadoda
 11/5/2020 2:17:14 PM

Quant Time: Nov 05 04:50:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1035969	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.14	114	4671901	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	4272015	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	3040571	10.02	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	26799	0.590	ppbv	98
13) Ethanol	3.60	45	7183m	0.729	ppbv	
15) Acetone	3.83	43	202569	1.829	ppbv	89
20) Methylene Chloride	4.31	84	32938	0.422	ppbv	96
26) Hexane	5.65	57	20994m	0.174	ppbv	
52) Tetrachloroethene	11.19	164	7046m	0.043	ppbv	
53) Toluene	9.77	91	45729m	0.109	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110420\
Data File : VL035841.D
Acq On : 4 Nov 2020 20:06
Operator : SY/AP
Sample : L4610-06DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV6DDL

Manual Integrations
APPROVED

MMDadoda
11/5/2020 2:17:14 PM

Quant Time: Nov 05 04:50:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

