

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110420\
 Data File : VL035849.D
 Acq On : 5 Nov 2020 1:59
 Operator : SY/AP
 Sample : L4220-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA3DL

Manual Integrations
 APPROVED

MMDadoda
 11/5/2020 2:17:56 PM

Quant Time: Nov 05 05:23:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	845949	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.14	114	3458363	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	3179699	10.00	ppbv	-0.04

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2227210	9.86	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.00	41	6033m	0.163	ppbv	
13) Ethanol	3.60	45	7855	0.976	ppbv	95
15) Acetone	3.83	43	38873m	0.430	ppbv	
20) Methylene Chloride	4.32	84	57638	0.905	ppbv	95
26) Hexane	5.66	57	33152	0.337	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110420\
Data File : VL035849.D
Acq On : 5 Nov 2020 1:59
Operator : SY/AP
Sample : L4220-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA3DL

Manual Integrations
APPROVED

MMDadoda
11/5/2020 2:17:56 PM

Quant Time: Nov 05 05:23:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

