

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110425\
 Data File : VL043152.D
 Acq On : 04 Nov 2025 12:59
 Operator : SY/MD
 Sample : Q3530-13 0.03 LOD
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT4-2025

Quant Time: Nov 05 06:51:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Oct 15 02:12:58 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 11/05/2025
 Supervised By : Mahesh Dadoda 11/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	113638	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	346779	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	298519	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	225526	9.720	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.583	62	255	0.041	ppbv #	64
32) 1,1,1-Trichloroethane	3.366	97	866m	0.033	ppbv	
35) Carbon Tetrachloride	3.774	117	960m	0.039	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110425\
Data File : VL043152.D
Acq On : 04 Nov 2025 12:59
Operator : SY/MD
Sample : Q3530-13 0.03 LOD
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT4-2025

Quant Time: Nov 05 06:51:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Oct 15 02:12:58 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/05/2025
Supervised By :Mahesh Dadoda 11/07/2025

