

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110520\
 Data File : VL035878.D
 Acq On : 6 Nov 2020 3:06
 Operator : SY/AP
 Sample : L4618-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-2DL

Manual Integrations
 APPROVED

MMDadoda
 11/9/2020 10:48:14 AM

Quant Time: Nov 08 18:47:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 0
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1027611	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	3773460	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	3484916	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2450482	9.90	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.00%

Target Compounds					Ovalue
9) Propene	3.00	41	108389m	2.406	ppbv
15) Acetone	3.83	43	1900148	17.296	ppbv 98
53) Toluene	9.78	91	1172975	3.465	ppbv 98
59) m/p-Xylene	12.95	91	54829m	0.153	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110520\
Data File : VL035878.D
Acq On : 6 Nov 2020 3:06
Operator : SY/AP
Sample : L4618-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-2DL

Manual Integrations
APPROVED

MMDadoda
11/9/2020 10:48:14 AM

Quant Time: Nov 08 18:47:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

