

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL110520\
 Data File : VL035884.D
 Acq On : 6 Nov 2020 6:56
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 11/9/2020 10:48:25 AM

Quant Time: Nov 06 08:13:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 0
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	978118	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	3776471	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	3436723	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2381556	9.75	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.86	43	13307m	0.127	ppbv	
20) Methylene Chloride	4.33	84	7979m	0.108	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL110520\
Data File : VL035884.D
Acq On : 6 Nov 2020 6:56
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
11/9/2020 10:48:25 AM

Quant Time: Nov 06 08:13:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

