

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110625\
 Data File : VL043168.D
 Acq On : 06 Nov 2025 15:27
 Operator : SY/MD
 Sample : Q3556-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 131-SG-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/07/2025
 Supervised By :Mahesh Dadoda 11/10/2025

Quant Time: Nov 07 00:23:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025
 QLast Update : Wed Oct 15 02:12:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	112949	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	318311	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	264373	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	205406	9.996	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.000%
Target Compounds						
					Qvalue	
13) Ethanol	1.722	45	9199	3.228	ppbv	95
15) Acetone	1.835	43	20709	1.360	ppbv	98
20) Methylene Chloride	2.062	84	2190	0.364	ppbv #	92
29) Chloroform	2.848	83	6795	0.270	ppbv	87
34) 2-Butanone	2.557	43	3680	0.172	ppbv	95
52) Tetrachloroethene	8.218	164	1171m	0.108	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110625\
Data File : VL043168.D
Acq On : 06 Nov 2025 15:27
Operator : SY/MD
Sample : Q3556-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
131-SG-1DL

Quant Time: Nov 07 00:23:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Oct 15 02:12:58 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/07/2025
Supervised By :Mahesh Dadoda 11/10/2025

