

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110625\
 Data File : VL043169.D
 Acq On : 06 Nov 2025 15:59
 Operator : SY/MD
 Sample : Q3556-04DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 131-SG-2DL

Quant Time: Nov 07 00:24:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Oct 15 02:12:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	110430	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.956	114	313872	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	260123	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	201260	9.954	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.500%
Target Compounds						
					Qvalue	
13) Ethanol	1.722	45	6455	2.316	ppbv	96
15) Acetone	1.836	43	23179	1.557	ppbv	91
19) Isopropyl Alcohol	1.891	45	1817m	0.196	ppbv	
20) Methylene Chloride	2.062	84	1799	0.306	ppbv	85
29) Chloroform	2.849	83	3542	0.144	ppbv #	78
34) 2-Butanone	2.564	43	4037	0.191	ppbv	97
36) Benzene	3.658	78	8773	0.297	ppbv #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110625\
Data File : VL043169.D
Acq On : 06 Nov 2025 15:59
Operator : SY/MD
Sample : Q3556-04DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
131-SG-2DL

Quant Time: Nov 07 00:24:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Oct 15 02:12:58 2025
Response via : Initial Calibration

