

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110625\
 Data File : VL043171.D
 Acq On : 06 Nov 2025 17:03
 Operator : SY/MD
 Sample : Q3557-05DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 141-SG-2DL

Quant Time: Nov 07 00:24:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Oct 15 02:12:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	111726	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	310352	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	260144	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	204363	10.107	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.502	85	1859	0.117	ppbv	97
13) Ethanol	1.722	45	3586	1.272	ppbv	96
15) Acetone	1.835	43	115569	7.674	ppbv	99
19) Isopropyl Alcohol	1.887	45	1703	0.182	ppbv #	1
20) Methylene Chloride	2.062	84	1868	0.314	ppbv #	83
34) 2-Butanone	2.560	43	5741	0.275	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110625\
Data File : VL043171.D
Acq On : 06 Nov 2025 17:03
Operator : SY/MD
Sample : Q3557-05DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
141-SG-2DL

Quant Time: Nov 07 00:24:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Oct 15 02:12:58 2025
Response via : Initial Calibration

