

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110722\
 Data File : VL039840.D
 Acq On : 7 Nov 2022 14:57
 Operator : SY/AP
 Sample : QC110322 CAN 10198
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC110322 CAN 10198

Quant Time: Nov 08 06:14:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Nov 03 09:27:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	1032770	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.652	114	2534091	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.459	117	2328531	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.770	95	1795023	10.299	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110722\
Data File : VL039840.D
Acq On : 7 Nov 2022 14:57
Operator : SY/AP
Sample : QC110322 CAN 10198
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC110322 CAN 10198

Quant Time: Nov 08 06:14:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Nov 03 09:27:12 2022
Response via : Initial Calibration

