

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110722\
 Data File : VL039841.D
 Acq On : 7 Nov 2022 15:56
 Operator : SY/AP
 Sample : N5409-01DL 2.5X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E062-1100DL

Quant Time: Nov 08 06:14:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Nov 03 09:27:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	1043545	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2574842	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2333355	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	1886489	10.802	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 108.000%	
Target Compounds						
						Qvalue
4) Chloromethane	2.829	50	418195	5.399	ppbv	98
9) Propene	2.710	41	734782	11.336	ppbv	98
10) Heptane	7.574	43	2024109	13.035	ppbv	100
30) Cyclohexane	6.600	84	672539	6.894	ppbv	98
36) Benzene	6.369	78	3056784	14.074	ppbv	96
39) 1,2-Dichloropropane	7.070	63	850544	10.051	ppbv	96
49) Bromoform	12.394	173	1147792	9.780	ppbv	98
57) Chlorobenzene	11.516	112	1158962	6.548	ppbv	94
59) m/p-Xylene	12.368	91	2590830	9.838	ppbv	92
60) o-Xylene	13.053	91	2842829	11.394	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110722\
Data File : VL039841.D
Acq On : 7 Nov 2022 15:56
Operator : SY/AP
Sample : N5409-01DL 2.5X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E062-1100DL

Quant Time: Nov 08 06:14:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Nov 03 09:27:12 2022
Response via : Initial Calibration

