

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110921\
 Data File : VL037987.D
 Acq On : 9 Nov 2021 20:07
 Operator : SY/AP
 Sample : M4555-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0U03DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/15/2021
 Supervised By :Mahesh Dadoda 11/17/2021

Quant Time: Nov 10 04:49:42 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Nov 9 04:49:42 2021
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.665	49	1506817	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.169	114	4228351	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.076	117	3575814	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.407	95	2501661	9.625	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.200%
Target Compounds						Qvalue
9) Propene	3.028	41	26132	0.260	ppbv	85
13) Ethanol	3.607	45	782867	77.159	ppbv	89
15) Acetone	3.861	43	198524	1.594	ppbv	96
19) Isopropyl Alcohol	3.976	45	47622m	0.598	ppbv	
20) Methylene Chloride	4.343	84	128635m	1.882	ppbv	
26) Hexane	5.684	57	87374m	0.499	ppbv	
29) Chloroform	5.745	83	55507m	0.212	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110921\
Data File : VL037987.D
Acq On : 9 Nov 2021 20:07
Operator : SY/AP
Sample : M4555-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0U03DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/15/2021
Supervised By : Mahesh Dadoda 11/17/2021

Quant Time: Nov 10 04:49:42 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

