

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110921\
 Data File : VL037996.D
 Acq On : 10 Nov 2021 2:01
 Operator : SY/AP
 Sample : M4555-07DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 C0U21DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/15/2021
 Supervised By :Mahesh Dadoda 11/17/2021

Quant Time: Nov 10 04:51:26 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Nov 09 04:51:26 2021
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.668	49	1267438	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.176	114	3921681	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.082	117	3316402	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.407	95	2237360	9.281	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.800%
Target Compounds						
13) Ethanol	3.633	45	34309m	4.020	ppbv	Qvalue
15) Acetone	3.880	43	50842m	0.485	ppbv	
20) Methylene Chloride	4.353	84	49447m	0.860	ppbv	
26) Hexane	5.690	57	24120m	0.164	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110921\
Data File : VL037996.D
Acq On : 10 Nov 2021 2:01
Operator : SY/AP
Sample : M4555-07DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
C0U21DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/15/2021
Supervised By :Mahesh Dadoda 11/17/2021

Quant Time: Nov 10 04:51:26 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

