

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111220\
 Data File : VL035916.D
 Acq On : 12 Nov 2020 16:43
 Operator : SY/AP
 Sample : L4720-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CVS-2DL

Manual Integrations
 APPROVED

MMDadoda
 11/16/2020 2:28:16 PM

Quant Time: Nov 13 03:01:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1223298	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	4170865	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3881567	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2844315	10.31	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	11734m	0.219	ppbv	
13) Ethanol	3.61	45	4129475	355.003	ppbv	96
15) Acetone	3.85	43	137454	1.051	ppbv	89
19) Isopropyl Alcohol	3.98	45	1718672	21.385	ppbv	100
20) Methylene Chloride	4.34	84	83476	0.907	ppbv	97
26) Hexane	5.68	57	43823	0.308	ppbv #	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111220\
Data File : VL035916.D
Acq On : 12 Nov 2020 16:43
Operator : SY/AP
Sample : L4720-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CVS-2DL

Manual Integrations
APPROVED

MMDadoda
11/16/2020 2:28:16 PM

Quant Time: Nov 13 03:01:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

