

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111220\
 Data File : VL035922.D
 Acq On : 12 Nov 2020 20:39
 Operator : SY/AP
 Sample : L4720-05DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CVS-5DL

Manual Integrations
 APPROVED

MMDadoda
 11/16/2020 2:28:26 PM

Quant Time: Nov 13 03:21:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1105853	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3891338	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3650239	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2640954	10.18	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	9660m	0.199	ppbv	
13) Ethanol	3.61	45	46806	4.451	ppbv	97
15) Acetone	3.86	43	83934	0.710	ppbv #	85
19) Isopropyl Alcohol	3.98	45	19168m	0.264	ppbv	
20) Methylene Chloride	4.34	84	148611	1.786	ppbv	97
26) Hexane	5.68	57	100892	0.785	ppbv #	45
53) Toluene	9.81	91	95957	0.275	ppbv	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111220\
Data File : VL035922.D
Acq On : 12 Nov 2020 20:39
Operator : SY/AP
Sample : L4720-05DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CVS-5DL

Manual Integrations
APPROVED

MMDadoda
11/16/2020 2:28:26 PM

Quant Time: Nov 13 03:21:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

