

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111220\
 Data File : VL035934.D
 Acq On : 13 Nov 2020 4:30
 Operator : SY/AP
 Sample : L4721-06
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LA-6

Quant Time: Nov 13 07:38:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 04 02:40:06 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1051359	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3624588	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	3302462	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2454149	10.46	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	5358m	0.536	ppbv	
15) Acetone	3.87	43	16358	0.146	ppbv #	16
20) Methylene Chloride	4.34	84	8398	0.106	ppbv #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111220\
Data File : VL035934.D
Acq On : 13 Nov 2020 4:30
Operator : SY/AP
Sample : L4721-06
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LA-6

Quant Time: Nov 13 07:38:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

