

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111323\
 Data File : VL040840.D
 Acq On : 13 Nov 2023 13:55
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 11/15/2023
 Supervised By :Semsettin Yesilyurt 11/15/2023

Quant Time: Nov 13 15:33:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Mon Nov 13 15:18:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.188	49	1134164	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.645	114	2904674	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.449	117	2574303	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	2054936	10.724	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.200%
Target Compounds						
						Qvalue
5) Vinyl Chloride	2.931	62	2021m	0.030	ppbv	
32) 1,1,1-Trichloroethane	5.995	97	4618m	0.025	ppbv	
35) Carbon Tetrachloride	6.484	117	4782m	0.029	ppbv	
38) Trichloroethene	7.278	130	3318m	0.034	ppbv	
52) Tetrachloroethene	10.593	164	2432m	0.027	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.034	83	6096m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111323\
Data File : VL040840.D
Acq On : 13 Nov 2023 13:55
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Nov 13 15:33:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Nov 13 15:18:50 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 11/15/2023
Supervised By : Semsettin Yesilyurt 11/15/2023

