

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111418\
 Data File : VL032746.D
 Acq On : 14 Nov 2018 4:59
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 11/16/2018 1:29:47 PM

Quant Time: Nov 14 06:04:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 14 06:04:57 2018
 QLast Update : Wed Nov 14 05:38:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1632080	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	4048701	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3681284	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2740254	9.88	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.30	62	8227	0.106	ppbv	# 86
32) 1,1,1-Trichloroethane	6.62	97	29647	0.109	ppbv	97
35) Carbon Tetrachloride	7.14	117	32975m	0.114	ppbv	
38) Trichloroethene	7.96	130	11562m	0.094	ppbv	
52) Tetrachloroethene	11.38	164	10968m	0.096	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.85	83	32365m	0.111	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111418\
Data File : VL032746.D
Acq On : 14 Nov 2018 4:59
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
11/16/2018 1:29:47 PM

Quant Time: Nov 14 06:04:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 05:38:10 2018
Response via : Initial Calibration

