

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111821\
 Data File : VL038051.D
 Acq On : 18 Nov 2021 11:55
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/25/2021
 Supervised By :Mahesh Dadoda 11/29/2021

Quant Time: Nov 18 12:28:40 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 10:39:10 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	934179	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2697914	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	2322240	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1675810	9.99	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds					Ovalue
5) Vinyl Chloride	3.27	62	2083m	0.025	ppbv
32) 1,1,1-Trichloroethane	6.51	97	5973m	0.024	ppbv
35) Carbon Tetrachloride	7.01	117	5990m	0.032	ppbv
38) Trichloroethene	7.83	130	3189m	0.031	ppbv
52) Tetrachloroethene	11.23	164	3696m	0.038	ppbv
63) 1,1,2,2-Tetrachloroethane	13.67	83	7171m	0.035	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111821\
Data File : VL038051.D
Acq On : 18 Nov 2021 11:55
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Quant Time: Nov 18 12:28:40 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 10:39:10 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 11/25/2021
Supervised By :Mahesh Dadoda 11/29/2021

