

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111821\
 Data File : VL038075.D
 Acq On : 19 Nov 2021 6:00
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/25/2021
 Supervised By :Mahesh Dadoda 11/29/2021

Quant Time: Nov 22 06:26:02 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	750377	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.16	114	2164144	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.07	117	1852806	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1328561	9.94	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.40%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Ovalue
52) Tetrachloroethene	11.20	164	8201m	0.103	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111821\
Data File : VL038075.D
Acq On : 19 Nov 2021 6:00
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Nov 22 06:26:02 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 11/25/2021
Supervised By :Mahesh Dadoda 11/29/2021

