

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111921\
Data File : VL038080.D
Acq On : 19 Nov 2021 11:36
Operator : SY/AP
Sample : VL1119ABL01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1119ABL01

Quant Time: Nov 22 06:51:42 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	776835	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2252370	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	1867405	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1387400	10.30	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.00%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111921\
Data File : VL038080.D
Acq On : 19 Nov 2021 11:36
Operator : SY/AP
Sample : VL1119ABL01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1119ABL01

Quant Time: Nov 22 06:51:42 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

