

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111921\
 Data File : VL038087.D
 Acq On : 19 Nov 2021 16:17
 Operator : SY/AP
 Sample : M4704-04DL2DUP 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 TRANS-SHOPDL2DUP

Quant Time: Nov 25 21:02:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	844678	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	2292911	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.06	117	1954042	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	1453064	10.31	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.00	41	81007	1.240	ppbv	99
15) Acetone	3.84	43	333442	4.553	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111921\
Data File : VL038087.D
Acq On : 19 Nov 2021 16:17
Operator : SY/AP
Sample : M4704-04DL2DUP 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
TRANS-SHOPDL2DUP

Quant Time: Nov 25 21:02:37 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

