

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112222\
 Data File : VL039929.D
 Acq On : 22 Nov 2022 23:27
 Operator : SY/AP
 Sample : N5767-05DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Quant Time: Nov 23 04:53:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Nov 22 03:45:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.172	49	962140	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.635	114	2647871	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.455	117	2320403	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	1832171	10.338	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.400%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.709	85	26599	0.193	ppbv	# 79
9) Propene	2.681	41	18650	0.298	ppbv	86
13) Ethanol	3.234	45	31175	6.529	ppbv	88
15) Acetone	3.459	43	96129	0.753	ppbv	98
20) Methylene Chloride	3.915	84	106899	1.729	ppbv	92
26) Hexane	5.192	57	73079	0.714	ppbv	# 40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112222\
Data File : VL039929.D
Acq On : 22 Nov 2022 23:27
Operator : SY/AP
Sample : N5767-05DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Quant Time: Nov 23 04:53:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Nov 22 03:45:31 2022
Response via : Initial Calibration

